

الأجهزة العلمية التي تحتاج إلى صيانة

القسم/الفسولوجي - كلية الطب البشري

اسم الجهاز والموديل	الشركة الموردة	العدد	الموقع	ملاحظات والأعطال التي تحصل عليها
Balance Mettler Model: AE ١٦٠ SN: ٣٨٦٠٠	Najm Al-Jazira	١	New Bldg. ١ st Flr. Preparation Rm. ٢٩	
Micro Osmometer Model: ٣M٠ SN: ٣٥٢٤٤A	Najm Al-Jazira	١	New Bldg. ١ st Flr. Lab Rm. ٢٨	Needs Calibration and Accessories.
Swelab Autocounter ٩١٠ Model: AC٩١٠ SN: ٠٢١٢ SN: ٠٢٠٣	Al-Gosaibi	٢	New Bldg. ١ st Flr. Lab Rm. ٢٨, ٣٢	Needs maintenance and calibration before the term starts because we can't keep it running because it will block the electrodes and they are very expensive to buy and replaced.
Pyroplate (Heater) Model: ١١٨٠-١ SN: ٠٤٩٢-٠٤٤٦	Najm Al-Jazira	١	New Bldg. ١ st Flr. Lab Rm. ٢٨	
Gallenamp Laboratory Centrifuge Model: CF٥٩٠ (No serial number)	Najm Al-Jazira	١	New Bldg. ١ st Flr. Lab Rm. ٢٨	
Gallenamp Water Bath Model: TM ٨٤٠ Mommert Model: W٢٧٠ (Both No serial number)	Najm Al-Jazira	٢	New Bldg. ١ st Flr. Lab Rm. ٢٨	
Aquatron Distillation Unit SN: ٤ S٠١٧٩٢ / Model: A٤٠٠٠	Najm Al-Jazira	١	New Bldg. ١ st Flr. Lab Rm. ٣٠	Needs cleaning because it get blocked.
Healthometer Model: ٤٠٢KGS SN:١٥٩٤٦٠ SN:١٥٩٤٦١ SN:١٥٩٤٧٣ SN:١٦٤٩٧٦		٤	New Bldg. ١ st Flr. Lab Rm. ٢٨, ٣٢	
Topcon Model: ACP٧M SN: ٨١٦٢٨٨ SN: ٨١٦٢٩٠	Al-Gosaibi	٢	New Bldg. ١ st Flr. Rm. ٢٧, Lab Rm. ٣٢	Lamp or bulb

اسم الجهاز والموديل	الشركة المورد	العدد	الموقع	ملاحظات والأعطال التي تحصل عليها
Student Spirometer Model: Student model Spirometer SN: K111865 SN: K111866 SN: K111867 SN: K111868	Al- Faras	4	New Bldg. 1st Flr. Lab Rm. 32	Pen and Tubings
Perimeter (No serial and Model)		3	New Bldg. 1st Flr. Lab Rm. 32	To Fix chin rest.
ZEISS Microscope (No serial and model)		13	New Bldg. 1st Flr. Lab Rm. 32	Bulb
Gowllands Ophthalmoscope Model: 300AM (No serial number)	Al- Faras	20	New Bldg. 1st Flr. Lab Rm. 32	Batteries
Microlab Spirometer Model: Microlab 3500 SN: 19063 SN: 19066 SN: 19065 SN: 19064	Najm Al-Jazira	4	New Bldg. 1st Flr. Lab Rm. 32	Thermal Paper
Recording Drum (Old machine, No model and serial number)	Najm Al-Jazira	2	New Bldg. 1st Flr. Lab Rm. 32	Paper & needs maintenance
Kymograph Bioscience Model: 1000 SN: 814 SN: 811 SN: 812	Najm Al-Jazira	3	New Bldg. 1st Flr. Lab Rm. 32	Paper & needs maintenance
Tanita Body Composition Analyzer Model: BF350 SN: 0212003 SN: 0211004	Al-Gosaibi	2	New Bldg. 1st Flr. Lab Rm. 28	
ECG Machine: *Cardiofax ECG Model: ECG-6801K SN: 04797 *Cardifax ECG Model: ECG-6011 SN: 20644	Najm Al-Jazira	2	New Bldg. 1st Flr. Lab Rm. 28	Accessories to be replaced Like chest leads.

اسم الجهاز والموديل	الشركة المورد	العدد	الموقع	ملاحظات والأعطال التي تحصل عليها
RIESTER Sphygmomanometer Model: CE ٠١٢٤٩ (No serial Number)	Najm Al-Jazira	١٥	New Bldg. ١ st Flr. Lab Rm. ٢٨	Needs to change the tubings and pipes
Easylyte Na/K Analyzer SN: ٢٦٢١٩ANKC SN: ١٩٧٧٤ ANK (No model number)	Al-Gosaibi	٢	New Bldg. ١ st Flr. Lab Rm. ٢٨, ٣٢	Needs maintenance and calibration before the term starts because we can't keep it running because it will block the electrodes and they are very expensive to buy and replaced.
Audiometer: * Siemens audiometer Model: SD٢١ SN: ٠٣٦٨٧٤ *Siemens T-٣١ audiometer Model: ٢٥٨٨١١٩ SN: ٢٤٨ * AMPLAID audiometer Model: ١٣١ SN: ٣٠٦٠٢	Najm Al-Jazira	٣	New Bldg. ١ st Flr. Lab Rm. ٣٢	Head phones to be replaced Yearly.
DSP Pure Tone Audiometer SN: ٠١٠٢٠٢٠٠٠٨DS SN: ٠١٠٢٠٢٠٠٠٩٣A SN: ٠١٠٢٠٢٠٠٠٩٠٤ (No model number)		٣	New Bldg. ١ st Flr. Lab Rm. ٣٢	Batteries
Swelab Mixer Model: ٤٤٠ SN: C٣٠٤٠ SN: C٣٠٢٤	Al-Gosaibi	٢	New Bldg. ١ st Flr. Lab Rm. ٢٨, ٣٢	
Sable system CA-١B CO٢ Analyzer SN: CAB٩٩٠١-٠٧ (No model number,old machine)	Najm Al- Jazira	١	New Bldg. ١ st Flr. Lab Rm. ٣٢	
Sable system PA-١B O٢ Analyzer SN: PAB٩٩٠١-٠٣E (No model number,old machine)	Najm Al- Jazira	١	New Bldg. ١ st Flr. Lab Rm. ٣٢	

اسم الجهاز والموديل	الشركة الموردة	العدد	الموقع	ملاحظات والأعطال التي تحصل عليها
Harvard Dry Gas Meter SN: ٨٧٠ (No model number) Dry Gas Meter SN: ٠٦١٧٥ (Old machine, No model number)	Najm Al-Jazira	٢	New Bldg. ١ st Flr. Lab Rm. ٣٢	Batteries
Micro Hematocrit Centrifuge SN: ٣٧٩٤٢٣ SN: ٥٧١٦٤ SN: ٥٧١٦٨ (Old machine ,No model number)	Najm Al-Jazira	٣	New Bldg. ١ st Flr. Lab Rm. ٢٨	
*NEC Model: VT٤٤٠K SN: ١٦٤٠٠٥٧NE *NEC Model: VT٤٥K SN: ٢٩٤٠٠٢٢NA		٢	New Bldg. ١ st Flr. Lab Rm. ٢٧	Bulbs